

MDCG 2024-13

Regulatory status of ethylene oxide (EtO) intended for the sterilisation of medical devices

October 2024

This document has been endorsed by the Medical Device Coordination Group (MDCG) established by Article 103 of Regulation (EU) 2017/745. The MDCG is composed of representatives of all Member States and it is chaired by a representative of the European Commission.

The document is not a European Commission document and it cannot be regarded as reflecting the official position of the European Commission. Any views expressed in this document are not legally binding and only the Court of Justice of the European Union can give binding interpretations of Union law.

1. Introduction

The regulatory status of ethylene oxide (EtO) for the sterilisation of single-use medical devices during the manufacturing process has been discussed in relation to the Review Programme under Regulation (EU) 528/2012 on biocidal products (BPR).

Ethylene oxide (EtO) is used to sterilise a wide range of medical products, such as medical devices, *in vitro* diagnostic medical devices (IVDs) and medicinal products, as well as products combining them. Sterilisation of products that are placed on the market in sterile condition is part of the manufacturing process and is usually carried out in large industrial units either by the manufacturers themselves or by external sterilisation service providers.

On a much smaller scale, EtO is also provided in cartridges for sterilisation of reusable medical devices, mainly in health institutions for use of such devices. Those EtO-containing cartridges, which are used for the supply of EtO to specific EtO sterilisers, are generally CE marked in accordance with the previous Directive 93/42/EEC or the current Regulation (EU) 2017/745 on medical devices (the 'MDR').

2. Regulatory status of EtO for sterilisation of medical devices

2.1. EtO sterilisation during the manufacturing process

EtO is a substance that is used in the manufacturing process of many medical devices and IVDs which are placed on the market in a sterile condition. As part of its Quality Management System (QMS) pursuant to Article 10(9) MDR and Article 10(8) IVDR, the relevant medical device manufacturers must address, among other matters, the sterilisation process and its validation, which is further specified e.g. in the harmonised standard *EN ISO 13485:2016 - Medical devices — Quality management systems — Requirements for regulatory purposes*.

In addition, several provisions in the MDR and IVDR specifically address the sterilisation of devices, in particular

- the general safety and performance requirements in Annex I MDR/IVDR (see sections 11.3-11.6 and 11.2-11.4 respectively), which are further specified in (harmonised) standards, such as
 - *EN ISO 11135:2014 + A1:2019: Sterilisation of health-care products - Ethylene oxide - Requirements for the development, validation and routine control of a sterilisation process for medical devices*,
 - *EN ISO 10993-7:2008 + AC:2009 + A1:2022: Biological evaluation of medical devices - Part 7: Ethylene oxide sterilization residuals*,
 - *EN 556-1:2024: Sterilization of medical devices - Requirements for medical devices to be designated "STERILE" - Part 1: Requirements for terminally sterilized medical devices*
- the technical documentation described in Annex II MDR/IVDR,
- the qualification requirements of notified body personnel laid down in Annex VII (sections 3.2.1 and 3.2.5) and
- the conformity assessment procedures laid down in Annex IX and XI MDR/IVDR (e.g. Annex IX, section 2.2 and 3.2 respectively).

As a consequence, EtO used for sterilisation during the manufacturing of medical devices and IVDs is used for a process that falls within the regulatory obligations imposed by the MDR and

IVDR. That means that this use of EtO is in scope of the MDR and IVDR even though EtO does not itself fall under the definitions of medical device or accessory for a medical device if used during the manufacturing process. Therefore, EtO falls outside the scope of the BPR, as its Article 2(2)(b) provides that the BPR does not apply in case of biocidal products or treated articles falling within the scope of the MDR or IVDR.

Where EtO is used for sterilisation during the manufacturing of medical devices and IVDs, its use is controlled as part of the sterilisation and validation processes of the manufacturer. These processes are subject to MDR/IVDR conformity assessment procedure in relation to the device being sterilised. EtO itself is not subject to additional MDR/IVDR conformity assessments.

2.2. EtO used by health institutions for sterilisation of medical devices before or after their use

EtO and the associated sterilising equipment can also be used by health institutions for sterilisation before first use of medical devices or after their use to enable their re-use. This is different from the use described in section 2.1.

In this case the EtO cartridges used for the sterilisation are in scope of the MDR by virtue of falling under the definition of medical device, or accessory for a medical device, and therefore are outside the scope of the BPR as per Article 2(2)(b) BPR.